



Labor Diagnostik Systeme GmbH

**Am Obstgarten 10
D-39326 Wolmirstedt / Germany**



The **CL-50 Plus** is a portable urine analyzer.

This device is a compact, cost effective system offering physician offices and smaller laboratories the reliability of instrument-based testing capability.

The unit tests levels of: pH, Nitrite, Specific Gravity, Blood, Glucose, Bilirubin, Urobilinogen, Ketone, Leucocytes, Protein, Calcium, Creatinine, Microalbumin.

Using dual-wavelength reflectance photometry, the unit two test modes:
Normal mode, 60 samples/hour.

The instrument uses cold light source facility with long life and high stability.

Bankverbindung

Commerzbank AG
Regionalfiliale München
Kontonummer: 21 21 43 200
Bankleitzahl: 700 400 41
IBAN: DE04 7004 0041 0212 1432 00
SWIFT: COBADEFF700

Geschäftsführer

Uwe Schipke

USt -IdNr.: DE272921521

HRB: 19210 Amtsgericht Stendal

Kontakt

Tel.: +49 (0) 39201 70 20 55
Fax: +49 (0) 39201 70 96 99
Email: info@labdiasys.de
www.labdiasys.de



Labor Diagnostik Systeme GmbH

**Am Obstgarten 10
D-39326 Wolmirstedt / Germany**

Technical Data

- Parameters: pH, Nitrite, Protein, Special Gravity, Blood, Glucose, Bilirubin, Urobilinogen, Ketone, Leucocytes Parameters, Calcium, Creatinine, Microalbumin.
- Principle: dual-wavelength reflectance photometry
- Throughput: Normal mode, 60 samples/hour
Continuous mode, 120 samples/hour
- Memory Capacity: 1000 results
- Reagent Strip: H10, H11, H13
- Display: LCD screen, operation information and test results
- Print Out: Internal thermal printer or you can hook up an optional external printer.
- Control function: Self-check, testing and malfunction-verification
- Serial Interface: RS232
- Weight: 2.0 kg
- Power: AC 110-240 V, 60/50 Hz
- Dimension: 290mm x 200mm x 100mm

Features & Functions

- Checking for ready-to-test strips automatically
- Internal calibration for sensitivity of strips
- External printer applicable & RS232 port to computer
- Barcode reader optional



Bankverbindung

Commerzbank AG
Regionalfiliale München
Kontonummer: 21 21 43 200
Bankleitzahl: 700 400 41
IBAN: DE04 7004 0041 0212 1432 00
SWIFT: COBADEFF700

Geschäftsführer

Uwe Schipke

USt -IdNr.: DE272921521

HRB: 19210 Amtsgericht Stendal

Kontakt

Tel.: +49 (0) 39201 70 20 55
Fax: +49 (0) 39201 70 96 99
Email: info@labdiasys.de
www.labdiasys.de