

Medical Grade Glycolide



Glycolide

Main features and benefits of Glycolide:

- Medical grade, certified according ISO13485
- REACH registered
- Purity: > 99.5 %
- Synonym: 1, 4-dioxane-2, 5-dione
- CAS number: 502-97-6

Glycolide is the dimer of glycolic acid and commonly used for the synthesis of glycolide containing homo- and copolymers. ITVP produces medical-grade glycolide of

highest purity via multistep purification. This allows reproducible synthesis of high molecular weight PGA-based polymers.

Glycolide

Description	ITVP-Glycolide is a medical grade monomer for ring opening polymerization. Glycolide, the dimer of glycolic acid, is a white, crystalline solid with high purity.		
Specification	Test	Method	Specification
	Appearance	Visual	With crystalline solid
	Identification	NMR	Conforms to reference
	Purity	NMR	> 99.5 %
	Melting range	DSC, 10 °C/min	82 – 87 °C
	Free Acid	NMR	< 0.038 wt%
	Water content	Titration	< 0.02 wt%
Heavy metal	---*	< 10 ppm	
Identification	Chemical name	1,4-dioxane-2,5-dione	
	Molecular formula	C ₄ H ₄ O ₄	
	Molecular weight	116.07 g/mol	
	CAS-No.	502-97-6	
Packaging	ITVP-Glycolide can be supplied in 1, 5 and 10 kg packages. Usual packaging consists of an inner PE and an outer bag of aluminum coated polyester-PE laminate. Products are shipped in additional boxes.		
Storage and Handling	ITVP recommends to store glycolide in the original packaging at low temperatures (< -15 °C). Allow glycolide to warm up to room temperature before opening the packaging. After opening the original packaging, please store glycolide under inert atmosphere at low temperatures.		

*Heavy metal content is measured from an external certificated laboratory.

ITV Denkendorf Produktservice GmbH (ITVP) is a subsidiary of the DITF (German Institutes of Textile and Fiber Research Denkendorf). ITVP is certified according ISO 13485 for the development and production of polymers, pre-products and devices for medical applications, but is not actively marketing medical devices for the final

consumer. The main focus lies on textile-based implants like e.g. surgical sutures and ligaments for wound treatment, meshes for hernia operations, vascular prostheses for blood vessel replacement, stents for use in the trachea and oesophagus, and membranes for treatment of burns.

ITV Denkendorf Produktservice GmbH

Contact: Dr. rer. nat. Sven Oberhoffner
T +49 (0)711 93 40-163
sven.oberhoffner@itvp-denkendorf.de

ITVP CEO: Prof. Dr.-Ing. Götz T. Gresser
T +49 (0)711 93 40-216
goetz.gresser@itvp-denkendorf.de

Koerschtalstraße 26 | D-73770 Denkendorf
www.itvp-denkendorf.de



Accredited laboratory by DAKKS according to DIN EN ISO/IEC 17025. The accreditation is valid only for the scope listed in the annex of the accreditation certificates D-PL-17474-01-00 respectively D-PL-17474-01-02.