

VivaChek™ 4-in-1

Multi-function Monitoring System



Glucose • Ketone • Uric Acid

VivaChek™ 4-in-1

Advanced Features



Bluetooth 4.2
(optional)



Tiny Sample Volume



Multi-function



Strip Ejector



Backlight
(optional)



Large Memories
(total)

Specification

	Ketone*	Uric Acid	Glucose (GDH-FAD)	Glucose (GOD)
Measurement Range	0.1 - 8.0 mmol/L	179 - 1190 µmol/L (3 - 20 mg/dL)	10 - 600 mg/dL (0.6 - 33.3 mmol/L)	10 - 600 mg/dL (0.6 - 33.3 mmol/L)
Testing Method	Electrochemical Biosensor	Electrochemical Biosensor	Electrochemical Biosensor	Electrochemical Biosensor
Coding	No Coding	No Coding	No Coding	No Coding
Test Time	10s	15s	5s	5s
Sample Volume	0.8 µL	2 µL	0.5 µL	0.5 µL
Sample Type	Fresh Whole Blood (capillary)	Fresh Whole Blood (capillary, venous)	Fresh Whole Blood (capillary, venous, arterial, neonatal)	Fresh Whole Blood (capillary, neonatal)
Operation Temperature Range	7.5 - 45 °C	10 - 40 °C	5 - 45 °C	5 - 45 °C
Hematocrit Range	20 - 65%	30 - 60%	0 - 70%	20 - 70%
Strip Shelf-life (unopened)	24 months	18 months	24 months	24 months
Strip Shelf-life (opened)	6 months	6 months	6 months	6 months
Strip Storage Temperature	2 - 30°C	4 - 38 °C	2 - 30 °C	2 - 30 °C
Humidity Exposure Limit for Strip	10 - 90% R.H.	R.H. ≤ 95%	10 - 90% R.H.	10 - 90% R.H.

*Ketone - the system only detects β-hydroxybutyric acid

VivaChek™
VivaChek Laboratories, Inc.
 913 N Market Street, Suite 200,
 Wilmington, DE 19801, USA
www.vivachek.com

Authorized distributor:

Number:1280022801

© 2019 VivaChek Laboratories, Inc.