


PRODUCT NAME	ELECTRICALLY OPERATED GYNECOLOGICAL CHAIR (TELESCOPIC MOTOR)
ORIGINAL PRODUCT CODE (P/N)	K015 ES
BRAND/MODEL OF PRODUCT	KENMAK K015 ES
PRODUCT PICTURE	
	
GYNECOLOGICAL-UROLOGICAL EXAMINATION TABLE	
Height adjustments (size, if available)	MAX 180 CM. SEAT PLATFORM POSITIONED BETWEEN MIN 70- MAX 92 CM
Backrest Length	90 - 53 CM
Seat Length	45 - 50 CM
Legrest Length	53 - 50 CM
Backrest Angle Adjustment (Angle, if available)	0- 80°
Trendelenburg – Reverse Trendelenburg Angle Adjustment (Angles, if available)	HORIZONTAL TRENDELENBURG MAX 20° - VIA ELECTRIC MOTOR IR HAND CONTROL UNIT
Height Range	70 - 92 CM
Height adjustment	VIA ELECTRIC MOTOR AND HAND UNIT- 220 MM
Movement structure (Manual, automatic)	AUTOMATIC - BACK MOVEMENT - HEIGHT MOVEMENT - TRENDELENBURG MOVEMENT
Control Mechanism (Foot or Hand Control Unit)	FOOT CONTROL UNIT (BRAND: DEWERT / LINAK / KENMAK) ADJUSTS THREE MOVEMENTS. THREE BUTTON SETS (IN TWO PAIRS) PROVIDE UP AND DOWN MOVEMENTS OF TRENDELENBURG, BACK AND HEIGHT. ALL THESE COMMANDS ADJUSTABLE VIA 6 BUTTONED DESK PANEL.
Number of Motors and Functions	ONE COLUMN, TWO 24 V POWERED DC MOTORS- TRENDELENBURG, BACK AND HEIGHT MOVEMENTS
Roll Placement Apparatus	N/A
Number of Behind Knee Support Apparatus and Specifications	JOINTED. 2 UNITS- MECHANICALLY UPWARDS AND DOWNWARDS ADJUSTABLE AND WITH HANDLE CLINCHING
Number of Arm Support Apparatus and Specifications	N/A
Number of Side Armrests and Specifications	VINILEX COATED ON THE STRAIN BAR- 2 UNITS
Waste Collection Basin	MADE OF CR 204 MATERIAL 32*17*15(DEPTH) - 1 UNIT
Serum Hanger	MADE OF 19*1,2 MM SIZE PIPE. 100 CM LONG. CHROME COATED. SINGLE UNIT. TWO HOOKS.
Handle	STRAIN BAR
Transporting chassis material	40*40*1.5 SIZE PROFILE PIPE- ELECTROSTATIC POWDER PAINT (STAINLESS)
Material of seat and rest platform	MADE OF 32 DENSITY GREY SPONGE COATED WITH ARTIFICIAL LEATHER ON 30*30*1.5 PROFILE PIPE. ANTISTATIC. DOES NOT HOLD LIQUID.
Wheels (diameter and brake system, if available)	N/A
Antistatic specification of wheels	N/A
Other specifications	N/A