

CERTIFICATE OF ANALYSIS

Product: FDP Clear High Control

Art. No: MRX192

Lot No: 22392

Exp. date: 2023 02

Assays performed on instrument: Sysmex CS2100i

Analysis	Specification	Result
Concentration FDP	20 – 40 µg/mL	30.4 µg/mL

Based on the above results and the quality assurance procedures of MediRox this lot is released for sale.

Issued by: Amanda Dyrefors
Date: 2021-05-04

Approved by: Veronica Regemar
Department: Quality Assurance
Date: 2021-05-06